

	<p>Am Dillhof 2 63863 Eschau - Germany</p> <p>Tel.: Int. +49 93 74 / 978 55 - 0 Fax: Int. +49 93 74 / 978 55 - 29</p> <p>Internet : http://www.isolab.de e-mail : sales@isolabgmbh.de</p>	<p><i>From the center of the German Laboratory- and Glass Industry - Eschau -</i></p>
---	--	---

9.02.2022

CERTIFICATE OF QUALITY ASSURANCE

Description:	Calibrated Density Bottle 100 ml	
Cat. No.:	S.023.02.100	
Serial No.:	204	
Visual appearance:	Within acceptability limits (According to EN ISO 2811-1 and ISO 3507).	
Calibration:	<p>According to ISO 4787, "Laboratory glassware – Volumetric glassware – Methods for use and testing of capacity". The method procedure is based upon a determination of volume of water contained by the vessel. This volume of water is based upon knowledge of its mass and its tabulated density at the reference temperature.</p> <p>This density bottle is calibrated according to DIN/ISO Class "A" tolerances.</p>	
Calculation:	Mass determined by gravimetric calibration: % 100	101,003 g
	Volume calculated acc. to ISO 4787: % 100	101,289 ml

The above mentioned density bottle is calibrated according to the mentioned standards and engraved the specific volume of use.

For and on behalf of

