



江苏省计量科学研究院

JIANGSU INSTITUTE OF METROLOGY

检定证书

Verification Certificate

证书编号:



Certificate No.

M2019-0105376

送检单位

Applicant

计量器具名称

Name of instrument

型号/规格

Type/Specification

出厂编号

Serial No.

制造单位

Manufacturer

检定依据

Verification Regulation

检定结论

Conclusion

量筒

见后

(自编号: 见后)

JJG 196-2006《常用玻璃量器检定规程》

合格



批准人

Approved by

胡强

核验员

Checked by

姚柠

检定员

Verified by

张晓晖

检定日期

Date of Verification

2019年

Year

11月

Month

18日

Day

有效期至

Valid to

2022年

Year

11月

Month

17日

Day

计量检定机构授权证书号: (国)法计(2017)01022

Authorization Certificate No. (2017)01022

地址: 南京市栖霞区文澜路95号(总部)

Add. No. 95, Wulan Road, Qixia District, Nanjing (Headquarters)

电话: (025) 84636990

Tel

电子信箱: jsimguest@jsim.com.cn

E-mail

网址: www.jsim.com.cn

Website

传真: (025)84636972

Fax

邮编: 210023

Post Code





本院是国家法定计量检定机构, 计量授权证书号: (国)法计(2017)01022号。

This laboratory is a national legal metrological verification institute. Authorization certificate No. (2017)01022

本院出具的数据均可溯源至国家计量基准。

All data issued by the laboratory are traceable to national measurement standards.

本次检定所使用的计量标准:

Measurement standard used in the verification

名称 Name	测量范围 Measuring Range	准确度等级/最大允许误差/ 不确定度 Class of Accuracy/Maximum Permissible Errors/Uncertainty of Measurement	证书编号/有效期至 Certificate No./Valid to
玻璃量器检定装置	(0.01~2000)mL	MPE: $\pm(12-0.4) \cdot 10^{-4}$	(1988)省 018 号/22-07-04

本次检定所使用的主要计量标准器具:

Main measurement standards used in the verification

名称/编号 Name/Number	测量范围 Measuring Range	准确度等级/最大允许误差/ 不确定度 Class of Accuracy/Maximum Permissible Errors/Uncertainty of Measurement	证书编号/有效期至 Certificate No./Valid to
电子天平/11636633	0.1mg~210g	1级	M2019-0102912/20-11-10
电子天平/220116912384	1mg~1200g	1级	M2019-0102917/20-11-10



检定地点、环境条件:

Place and environmental conditions of the verification

地点: 南京市栖霞区文澜路 95 号 7218 室

温度: 18.3°C

相对湿度: --%

Place No.95,Wenlan Road,Qixia District,Nanjing Room 7218

Temperature

R. Humidity

注: 1、未经本院书面授权, 不得部分复制本证书。

Note This certificate shall not be reproduced except in full, without the written approval of the laboratory.

2、本证书的检定结果仅对所检计量器具有效。

The results are only responsible for the items verified.

3、本证书未加盖检定专用章无效。

This certificate will be invalid if it is issued without official stamp.



检定结果

Verification Results

器具名称: 量筒			
自编号	标称容量 (mL)	容量允差 (mL)	备注
4077, 4078	10	± 0.10	In
4079, 4080	250	± 1.0	In
标准温度: 20°C			
以下空白			



江苏计量
Jiangsu Metrology

