

Certificate of Analysis

Product Identification:

Product Description	3M™ Comply™ 00135LF Bowie-Dick Plus Test Pack
3M Catalog Number	00135LF
3M Identification Number	70-2007-6072-9
3M Lot Number	202501CN
Shelf Life	2 years (24 months) from date of manufacture
Date of Manufacture	2023-01-13
Expiration Date	2025-01-12
Country of Manufacture	USA

Test Data:

Test	Specification	Result
ISO Inadequate Vacuum Test, 270° F / 3.5 minutes	Pass	Pass

The above lot number(s) meets ISO 11140-5:2007 for Type 2 Indicators for Bowie and Dick Type Air Removal Tests. Testing has been found to comply with 3M's quality performance specifications applicable at time of manufacture.

Prepared by:



QA representative:

Date: 2023-01-13