

	<p>D 97877 Wertheim - Germany Bahnhofstrasse 10 Tel.: Int. 0049-9342-912355 Fax: Int. 0049-9342-912357</p> <p>Internet : http://www.isolab.de e-mail : sales@isolabgmbh.de</p>	<p><i>From the center of the German Laboratory- and Glass Industry - Wertheim -</i></p>
---	---	---

26.11.2019

CERTIFICATE OF QUALITY ASSURANCE

Description:	Calibrated Density Bottle 50 ml				
Cat. No.:	S.023.02.050				
Serial No.:	365				
Visual appearance:	Within acceptability limits (According to DIN 12797 and ISO 3507).				
Calibration:	<p>According to ISO 4787, "Laboratory glassware – Volumetric glassware – Methods for use and testing of capacity". The method procedure is based upon a determination of volume of water contained by the vessel. This volume of water is based upon knowledge of its mass and its tabulated density at the reference temperature.</p> <p>This density bottle is calibrated according to DIN/ISO Class "A" tolerances.</p>				
Calculation:	<table style="width: 100%; border: none;"> <tr> <td style="padding-right: 20px;">Mass determined by gravimetric calibration: % 100</td> <td style="text-align: right;">50,208 g</td> </tr> <tr> <td>Volume calculated acc. to ISO 4787: % 100</td> <td style="text-align: right;">50,351 ml</td> </tr> </table>	Mass determined by gravimetric calibration: % 100	50,208 g	Volume calculated acc. to ISO 4787: % 100	50,351 ml
Mass determined by gravimetric calibration: % 100	50,208 g				
Volume calculated acc. to ISO 4787: % 100	50,351 ml				

The above mentioned density bottle is calibrated according to the mentioned standards and engraved the specific volume of use.

For and on behalf of

