



# 江苏省计量科学研究所

JIANGSU INSTITUTE OF METROLOGY

## 检定证书

Verification Certificate

证书编号: M2022-0092393  
Certificate No.

送检单位

Applicant

计量器具名称

Name of Instrument

型号/规格

Type/Specification

出厂编号

Serial No.

制造单位

Manufacturer

检定依据

Verification Regulation

检定结论

Conclusion

座式滴定管

5.0mL

(自编号: 见后)

JJG 196-2006 《常用玻璃量器检定规程》

合格 (A 级)



批准人

Approved by

陈力

核验员

Checked by

姚村

检定员

Verified by

张晓晔

检定日期

Date of Verification

2022 年

Year

9 月

Month

19 日

Day

有效期至

Valid to

2025 年

Year

9 月

Month

18 日

Day



地址: 南京市栖霞区文澜路 95 号 (总部)

Add: No.95, WenLan Lu, QiXia District, Nanjing (Headquarter)

电话: (025) 84636990

Tel

网址: www.jsim.com.cn

Website

传真: (025)84636972

Fax

电子信箱: jsimguest@jsim.com.cn

E-mail

邮编: 210023

Post Code



本院是国家法定计量检定机构, 计量授权证书号: (国) 法计 (2017) 01022 号。

This laboratory is a national legal metrological verification institute. Authorization certificate No. (2017)01022.

本证书中的检定结果均可溯源至国际单位制(SI)单位/社会公用计量标准。具体溯源链信息见附页。

The verification results in the certificate can be traceable to international system of units (SI) / social public measurement standards. See the attached page for specific information of traceability chain.

本次检定所使用的计量标准:

Measurement standard used in the verification

名称 Name	测量范围 Measuring Range	准确度等级/最大允许误差/ 不确定度 Class of Accuracy/Maximum Permissible Errors/ Uncertainty of Measurement	证书编号/有效期至 Certificate No./Valid to
玻璃量器检定装置	(0.01~2000)mL	MPE: $\pm(12 \sim 0.4) \times 10^{-4}$	[1988]省量标省法证字第 018号/27-06-12

本次检定所使用的主要计量标准器具:

Main measurement standards used in the verification

名称/编号 Name/Number	测量范围 Measuring Range	准确度等级/最大允许误 差/不确定度 Class of Accuracy/Maximum Permissible Errors/ Uncertainty of Measurement	溯源机构名称 Name of traceability institution	证书编号/有效期至 Certificate No./Valid to
电子天平/11636633	0.1mg~210g	I级	本院	M2021-0096692 /22-10-12

其他检定信息:

Other information of the verification:

地点: 南京市栖霞区文澜路 95 号 8203 室

Place No.95,Wenlan Road,Qixia District,Nanjing Room 8203

温度: 21.3°C

Temperature

相对湿度: --%

R. Humidity

到样日期: 2022 年 09 月 14 日

Date of sample received Year Month Day

注: 1、未经本院书面授权, 不得部分复制本证书。

Note This certificate shall not be reproduced except in full, without the written approval of the laboratory.

2、本证书的检定结果仅对所检计量器具有效。

The results are only responsible for the items verified.

3、本证书未加盖检定专用章无效。

This certificate will be invalid if it is issued without official stamp.



江苏计量  
Jiangsu Metrology

证书编号: M2022-0092393

Certificate No.

第 3 页 共 4 页

Page of

## 检定结果

Verification Results

自编号	标称容量 (mL)	级别	容量允差 (mL)	备注
1625	5.0	A	$\pm 0.010$	Ex
1626	5.0	A	$\pm 0.010$	Ex
标准温度: 20℃				

以下空白/Following blank



江苏计量  
Jiangsu Metrology





附页

Attached page

### 标准器溯源链信息

Traceability chain information for standard instruments

序号 No.	标准器名称 Name of standard instruments	溯源单位 Traceability unit	证书号/有效期 Certificate No./Valid to	溯源中所使用 标准器序号 Standard instruments No. used in traceability
1	电子天平	本院	M2021-0096692/22-10-12	1.1
1.1	砝码	本院	M2021-0049635,M2021-0074684/22-05-31	/



江苏计量  
Jiangsu Metrology

